



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **MLM Medical Labs**, using **Tosoh G8** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#17**

The system evaluated was:

|  |   |                               |
|--|---|-------------------------------|
| Instrument:<br><b>G8</b>                                   | Calibrator Lot:<br><b>GB2020</b>                    | Column Lot:<br><b>KC0889S</b> |
| Reagent Lot:<br><b>S8-103K, S8-203K, S8-302K, HW089D/B</b> | Calibrator Assigned Values:<br><b>5.89%, 10.94%</b> |                               |

Date of Certification: May 1, 2024

Certification Expires: May 1, 2025

A handwritten signature in black ink, appearing to read 'David Sacks', is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Randie R Little PhD', is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'C. J. [unclear]', is written over a horizontal line.

SRL director/ supervisor